



University of Kentucky
UKnowledge

Biomedical Engineering Faculty Publications

Biomedical Engineering

11-30-2013

Diffuse Correlation Spectroscopy (DCS) for Assessment of Tissue Blood Flow in Skeletal Muscle: Recent Progress

Yu Shang

University of Kentucky, yu.shang@uky.edu

Katelyn Gurley

University of Kentucky, Katelyn.Gurley@uky.edu

Guoqiang Yu

University of Kentucky, guoqiang.yu@uky.edu

Follow this and additional works at: https://uknowledge.uky.edu/cbme_facpub



Part of the [Biomedical Engineering and Bioengineering Commons](#)

Right click to open a feedback form in a new tab to let us know how this document benefits you.

Repository Citation

Shang, Yu; Gurley, Katelyn; and Yu, Guoqiang, "Diffuse Correlation Spectroscopy (DCS) for Assessment of Tissue Blood Flow in Skeletal Muscle: Recent Progress" (2013). *Biomedical Engineering Faculty Publications*. 4.

https://uknowledge.uky.edu/cbme_facpub/4

This Review is brought to you for free and open access by the Biomedical Engineering at UKnowledge. It has been accepted for inclusion in Biomedical Engineering Faculty Publications by an authorized administrator of UKnowledge. For more information, please contact UKnowledge@lsv.uky.edu.

Diffuse Correlation Spectroscopy (DCS) for Assessment of Tissue Blood Flow in Skeletal Muscle: Recent Progress

Digital Object Identifier (DOI)

<http://dx.doi.org/10.4172/2161-0940.1000128>

Notes/Citation Information

Published in *Anatomy & Physiology*, v. 3, issue 2, article 128, p. 1-4. © 2013 Shang Yu, et al. This is an open-access article distributed under the terms of the Creative Commons Attribution License, which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited.

Diffuse Correlation Spectroscopy (DCS) for Assessment of Tissue Blood Flow in Skeletal Muscle: Recent Progress

Yu Shang¹, Katelyn Gurley^{1,2} and Guoqiang Yu^{1*}

¹Department of Biomedical Engineering, University of Kentucky, Lexington, KY 40506, USA

²Department of Neurological Surgery, University of Louisville, Louisville, KY 40202, USA

Abstract

Near-infrared diffuse correlation spectroscopy (DCS) is an emerging technology for monitoring blood flow in various tissues. This article reviews the recent progress of DCS for the assessment of skeletal muscle blood flow, including the developments in technology allowing use during dynamic exercise and muscular electrical stimulation, the utilization for diagnosis of muscle vascular diseases, and the applications for evaluating treatment effects. The limitations of current DCS studies and future perspective are finally discussed.

Keywords: Diffuse Correlation Spectroscopy (DCS); Skeletal muscle; Blood flow; Motion artifact; Gating algorithm; Peripheral Arterial Disease (PAD)

Overview

Skeletal muscles comprise approximately 40% of total body mass, facilitating posture and locomotion and utilizing up to 85% total cardiac output [1]. Many vascular diseases, such as peripheral arterial disease (PAD), cause a decline in skeletal muscle blood flow [2-4], impairing the delivery of oxygen and other nutrients to the tissue. Correspondingly, many therapies have attempted to improve microcirculation and oxygen consumption in muscles. Monitoring of skeletal muscle blood flow and oxygen consumption rate is essential to provide insight into muscle pathophysiology and may help determine diagnosis and efficacy of treatment.

Blood flow in skeletal muscle has been quantified by a number of noninvasive methods. Conventional venous occlusion plethysmography has been employed in muscle perfusion investigations for many years [5,6]. This method, however, can be used only in the static state, since the occlusions interrupt muscle blood flow. Arterial-spin-labeled MRI (ASL-MRI) [7] and positron emission tomography (PET) [8] are also capable of monitoring microvasculature blood flow, but these methods require expensive and cumbersome equipment not available to all clinics or laboratories. Furthermore, these technologies are highly sensitive to motion artifacts which distort signals from the exercising muscle [9].

Oxygen consumption rate ($\dot{V}O_2$) has been quantified by a variety of techniques. Spirometry measures whole-body $\dot{V}O_2$, but cannot provide information about local muscle tissues [10]. Blood sampling is an established technique, but is invasive and remains regional in scope as the samples are taken from major vessels connected to groups of muscles [11,12]. The 31-phosphorous magnetic resonance spectroscopy (31P-MRS) can quantify local muscle $\dot{V}O_2$ [13], but requires expensive and cumbersome equipment which limits use in exercise studies. Additionally, ³¹P-MRS measurements have poor temporal resolution and limited sensitivity, inhibiting the ability to quantify rapidly changing dynamics.

A novel technique known as near-infrared (NIR) diffuse correlation spectroscopy (DCS) has been recently developed that enables quantifying relative changes in microvascular blood flow (rBF). DCS uses coherent near-infrared light to penetrate deep tissues and measures speckle fluctuations of the diffuse light that are sensitive

to the motions of red blood cells in tissues [3,14,15]. DCS provides a portable, noninvasive, and inexpensive alternative for microvascular blood flow measurements and has been validated against other standards, including power spectral Doppler ultrasound [16], Doppler ultrasound [17,18], laser Doppler flowmetry [19,20], Xenon computed tomography [21], fluorescent microsphere flow measurement [22], and ASL-MRI [23]. The DCS technology has been extensively introduced into various tissues, including brain [18,21,24-29], tumor [30-33] and skeletal muscle [2,3,23,34-42]. The applications of DCS in brain and tumor have been previously reviewed [43-46]. Interested readers are encouraged to read these publications for details. This review paper will focus on introducing some recent progress of DCS in the study of skeletal muscles.

Principle of DCS Technology

Historically, the motion of red blood cells in superficial tissues (<1 mm) was quantified by detecting light fluctuations within a single speckle area on tissue surface with a single scattering theory [47-49]. In thick/deep tissues such as skeletal muscles, the photons experience multiple scatterings with varied pathlengths. DCS is a theory that accounts for multiple scatterings of photons and quantifies the motion of red blood cells (i.e., the blood flow) in thick/deep tissues [14,50-52].

The key components of a DCS flow-meter include a long-coherence length laser at NIR range, a single-photon-counting avalanche photodiode (APD) detector and an autocorrelator board. Other components, such as source/detector fibers, computer, and A/D board, are used to couple light in/out of tissue or control/record optical data. For tissue blood flow measurement, the laser placed on the tissue surface (e.g., skin) launches long-coherence NIR light into the tissue via a multiple-mode source fiber, and the light transported/

***Corresponding author:** Guoqiang Yu, Department of Biomedical Engineering, University of Kentucky, Lexington, KY 40506-0070, USA, Tel: (859) 257-9110; E-mail: guoqiang.yu@uky.edu

Received October 15, 2013; **Accepted** November 28, 2013; **Published** November 30, 2013

Citation: Shang Y, Gurley K, Yu G, (2013) Diffuse Correlation Spectroscopy (DCS) for Assessment of Tissue Blood Flow in Skeletal Muscle: Recent Progress. *Anat Physiol* 3: 128. doi:[10.4172/2161-0940.1000128](https://doi.org/10.4172/2161-0940.1000128)

Copyright: © 2013 Shang Yu, et al. This is an open-access article distributed under the terms of the Creative Commons Attribution License, which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited.

scattered through the tissue was collected by a single-mode (or a few-mode) detector fiber placed millimeters to centimeters away from the source fiber. The detected light is then delivered via the detector fiber to APD detector, where the count of photons per unit time (i.e., light intensity) is recorded. The temporal fluctuation of light intensity in a single speckle area of tissue collected by the detector fiber is associated with the motion of moving scatterers (primarily red blood cells in microvasculature) and can be quantified by calculating the decay of light intensity autocorrelation function using the autocorrelator. The electric field temporal autocorrelation function of light is determined from the normalized light intensity autocorrelation function, and it obeys the correlation diffusion equation in highly scattering media [14,15]. By fitting the electric field autocorrelation curve to an analytical solution of correlation diffusion equation and assuming a particular flow model, the blood flow index is yielded [19,34]. Brownian diffuse motion is found to be the best flow model to fit the DCS autocorrelation curve, and the blood flow index is represented by αD_B [19,52]. Here D_B is the effective diffusive coefficient and α is the ratio of light scattering events from moving red blood cells to total scatters. For human skeletal muscles, a semi-infinite geometry is often assumed to obtain the analytical solution of correlation diffusion equation [2,19,39,41]. Because the absolute value of blood flow index (αD_B) is affected by many factors (e.g., the coupling between fiber and tissue surface) and varies across subjects, most of studies only reported the relative change in blood flow (rBF) during or post physiological manipulation as compared to the baseline (before physiological manipulation) [3,4,20,23,27,28,34,35,39,53].

Technology Development: Gating Algorithm and Absolute Measurement

Exercise is a commonly used protocol to evaluate the working capacity of skeletal muscle as well as its circulatory and metabolic functions. However, measurement of blood flow during exercise is a difficult task, as flow measurement is susceptible to muscle fiber motions during exercise. Previous study has demonstrated that the muscle fiber motions increase moving scattering in tissue, introducing artifacts into DCS measurements and leading to overestimation of blood flow changes during exercise [34]. To overcome this limitation, a novel algorithm was recently proposed to gate DCS data acquisition based on the muscle activity status [42]. Recordings from a dynamometer, such as hand/foot position, torque and velocity, were used to determine the muscle status (contraction vs. relaxation) during exercise. The DCS control program recorded data only at the time when muscle was relaxed.

This novel gating algorithm for minimizing the motion artifacts to blood flow measurements has been applied to a handgrip exercise study [42]. Figure 1 shows the blood flow index (i.e., αD_B) measured by DCS during handgrip exercise without (Figure 1a) and with (Figure 1b) use of the gating algorithm, respectively. Apparently, muscle fiber motions during exercise introduced remarkable artifacts into DCS blood flow measurement (Figure 1a). The gating algorithm significantly reduced the motion artifacts and enabled recording of the true blood flow signals during exercise (Figure 1b).

DCS is a valuable tool providing reliable quantity of relative changes in blood flow (i.e., rBF) rather than its absolute value (αD_B). To obtain an absolute measurement, a near-infrared spectroscopy (NIRS) oximeter was utilized in a calibration protocol to extract the absolute values of forearm blood flow and $\dot{V}O_2$ during a rest period (baseline) through venous and arterial occlusions of the upper arm




Figure 1: Blood flow index (αD_B) measured by DCS during exercise without (a) and with (b) use of gating algorithm. The dashed vertical lines indicate the beginning and ending of exercise. Courtesy of K. Gurley [54].


[42,54]. Briefly, three venous occlusions at 50 mmHg and one arterial occlusion at 240 mmHg were applied sequentially on the upper arm of the subject. Forearm flexor muscle oxygenation data, including oxy- and deoxy-hemoglobin concentrations ($[HbO_2]$ and $[Hb]$), total hemoglobin concentration (THC), and blood oxygen saturation (S_iO_2), were acquired using a commercial NIRS oximeter (Imagent, ISS Inc., IL). Baseline blood flow was derived as the average rate of initial increase in THC during the first 5 seconds of three venous occlusions. Baseline $\dot{V}O_2$ was calculated as the linear regression of the oxygen desaturation rate (i.e., $[HbO_2]$ - $[Hb]$) during the first 60 seconds of arterial occlusion. The absolute values of blood flow and $\dot{V}O_2$ throughout (i.e., pre, during and post) exercise were calculated by multiplying the rBF and $r\dot{V}O_2$ and their absolute baseline values measured by the occlusion protocols, respectively. Here $r\dot{V}O_2$ represents the relative change of $\dot{V}O_2$ that was calculated from relative changes of blood flow and oxygenation data measured by the DCS/NIRS devices [3,39,42,54]. Figure 2 shows the time course data of absolute blood flow (Figure 2a) and $\dot{V}O_2$ (Figure 2b) in forearm muscles of nine healthy subjects throughout a 3-minute handgrip exercise. Data were recorded by the DCS/NIRS devices with the gating algorithm and calibration protocols. The baseline values of blood flow and $\dot{V}O_2$ as well as their dynamic changes throughout exercises were found to agree with previous literature [55-57].

Very recently, the novel gating algorithm was adapted in a muscular electrical stimulation (ES) study to facilitate blood flow and oxygenation measurements during ES on quadriceps muscle [58]. According to the gating algorithm, DCS/NIRS data were only recorded in muscle fiber relaxation status (determined by output current from the muscle stimulator). Measurement of muscle blood flow/oxygenation responses to ES offers an objective standard to optimize ES treatment for diseases caused by poor muscle blood circulation and oxygenation (e.g., pressure ulcer).

As an advanced application, DCS was used in another ES study to extract the strain rate of contracting or relaxing muscle fibers through fitting the correlation curve to a unique physical model [40]. Fast- and slow-twitch muscle fibers were differentiated with this advanced technique, indicating the potential of using DCS to provide deeper insight into muscle physiology.

Diagnosis of Skeletal Muscle Diseases

DCS/NIRS has been used to assess muscle circulatory and metabolic function in diseased skeletal muscles [3,4,39]. A recent study investigated muscle rBF responses in healthy and PAD groups following treadmill and pedal exercises [4], showing significant differences between the two groups. Compared to the healthy subjects



who exhibited elevation in muscle rBF after exercise, the PAD patients maintained a relatively stable muscle flow throughout exercise, implying some vascular deficit in responding to the metabolic demands of exercise.

DCS/NIRS was also used to quantify muscle hemodynamic and metabolic responses in older women with fibromyalgia following isometric pedal exercise [39]. Compared to the age-matched healthy controls, patients with fibromyalgia exhibited lower oxygen extraction rate although their rBF values did not show a significant difference.

Evaluation of Surgical and Physical Treatment Effects

DCS technology was first applied in clinical surgery by monitoring muscle blood flow in patients with PAD during femoral arterial revascularization [2]. DCS demonstrated high sensitivity in detecting the instantaneous changes in rBF during surgical procedures as well as the rBF improvement immediately after revascularization. The acute improvements in calf muscle blood flow were associated with long-term significant improvements in symptoms and function. This pilot study corroborates potential of the diffuse optical methods for assessing the success of arterial revascularization.

In another application, DCS was used to evaluate the effect of massage on muscle microcirculation [38]. Immediate increase in calf muscle rBF was found following 8-min lower limb massage on young women, indicating the potential of DCS for assessing acute treatment effects of physical therapy.

Study Limitations and Future Perspective

Currently, the blood flow index (αD_b) in DCS technology is mostly calculated using the semi-infinite analytical solution [52,59,60], where the tissue is assumed to be homogenous. As a result, the yielding index reflects blood flow information not only from skeletal muscle, but also from skin and fat, as light also penetrates into these overlaying tissues. This phenomenon is well known as “partial volume effect” [27,28]. Realistically, blood flow responses from different tissues (skin, fat and muscle) differ remarkably [3], thus tissue heterogeneity should be taken

into account. One of future goals is to overcome the limitation of this “partial volume effect” and precisely extract the blood flow in skeletal muscles by combining multiple source-detector separation data with the DCS solution for heterogeneous tissues. Furthermore, the approach to control DCS data acquisition during exercise may be improved for further reducing motion artifacts. The current gating algorithm utilizes extrinsic signals, such as foot position/velocity/torque or ES current, to determine the muscle status during exercise or ES. Intrinsic signals detected directly from the muscle (e.g., by electromyography) may be employed to precisely gate DCS measurements. With further technology developments and more clinical applications, DCS is expected to become the technique of choice for diagnosis and treatment management of muscular diseases.

Acknowledgement

The authors are grateful for funding support from the National Institutes of Health (NIH) R01-CA149274, R21-AR062356 and UL-1RR033173. The content is solely the responsibility of the authors and does not necessarily represent the official views of the NIH.

References

1. De Blasi RA, Ferrari M, Natali A, Conti G, Mega A, et al. (1994) Noninvasive measurement of forearm blood flow and oxygen consumption by near-infrared spectroscopy. *J Appl Physiol* 76: 1388-1393.
2. Yu G, Shang Y, Zhao Y, Cheng R, Dong L, et al. (2011) Intraoperative evaluation of revascularization effect on ischemic muscle hemodynamics using near-infrared diffuse optical spectroscopies. *J Biomed Opt* 16: 027004.
3. Yu G, Durduran T, Lech G, Zhou C, Chance B, et al. (2005) Time-dependent blood flow and oxygenation in human skeletal muscles measured with noninvasive near-infrared diffuse optical spectroscopies. *J Biomed Opt* 10: 024027.
4. Mesquita RC, Putt M, Chandra M, Yu G, Xing X, et al. (2013) Diffuse optical characterization of an exercising patient group with peripheral artery disease. *J Biomed Opt* 18: 57007.
5. Whitney RJ (1953) The measurement of volume changes in human limbs. *J Physiol* 121: 1-27.
6. Engelke KA, Halliwill JR, Proctor DN, Dietz NM, Joyner MJ (1996) Contribution of nitric oxide and prostaglandins to reactive hyperemia in human forearm. *J Appl Physiol* 81: 1807-1814.
7. Schmitt P, Kotas M, Tobermann A, Haase A, Flentje M (2003) Quantitative tissue perfusion measurements in head and neck carcinoma patients before and during radiation therapy with a non-invasive MR imaging spin-labeling technique. *Radiother Oncol* 67: 27-34.
8. Baron JC (1999) Mapping the ischaemic penumbra with PET: implications for acute stroke treatment. *Cerebrovasc Dis* 9: 193-201.
9. Rådegran G (1999) Limb and skeletal muscle blood flow measurements at rest and during exercise in human subjects. *Proc Nutr Soc* 58: 887-898.
10. Smithies MN, Royston B, Makita K, Konieczko K, Nunn JF (1991) Comparison of oxygen consumption measurements: indirect calorimetry versus the reversed Fick method. *Crit Care Med* 19: 1401-1406.
11. Van Beekvelt MC, Colier WN, Wevers RA, Van Engelen BG (2001) Performance of near-infrared spectroscopy in measuring local O_2 consumption and blood flow in skeletal muscle. *J Appl Physiol* (1985) 90: 511-519.
12. Mottram RF (1955) The oxygen consumption of human skeletal muscle in vivo. *J Physiol* 128: 268-276.
13. Hamaoka T, Iwane H, Shimomitsu T, Katsumura T, Murase N, et al. (1996) Noninvasive measures of oxidative metabolism on working human muscles by near-infrared spectroscopy. *J Appl Physiol* (1985) 81: 1410-1417.
14. Boas DA, Campbell LE, Yodh AG (1995) Scattering and Imaging with Diffusing Temporal Field Correlations. *Phys Rev Lett* 75: 1855-1858.
15. Boas DA, Yodh AG (1997) Spatially varying dynamical properties of turbid media probed with diffusing temporal light correlation. *J Opt Soc Am A-Opt Image Sci Vis* 14: 192-215.
16. Yu G, Durduran T, Zhou C, Wang HW, Putt ME, et al. (2005) Noninvasive monitoring of murine tumor blood flow during and after photodynamic therapy

- provides early assessment of therapeutic efficacy. *Clin Cancer Res* 11: 3543-3552.
17. Buckley EM, Cook NM, Durduran T, Kim MN, Zhou C, et al. (2009) Cerebral hemodynamics in preterm infants during positional intervention measured with diffuse correlation spectroscopy and transcranial Doppler ultrasound. *Opt Express* 17: 12571-12581.
 18. Roche-Labarbe N, Carp SA, Surova A, Patel M, Boas DA, et al. (2010) Noninvasive optical measures of CBV, StO₂, CBF index, and rCMRO₂ in human premature neonates' brains in the first six weeks of life. *Hum Brain Mapp* 31: 341-352.
 19. Durduran T (2004) Non-invasive measurements of tissue hemodynamics with hybrid diffuse optical methods. Ph.D. dissertation 1-208, University of Pennsylvania, Philadelphia, USA.
 20. Shang Y, Chen L, Toborek M, Yu G (2011) Diffuse optical monitoring of repeated cerebral ischemia in mice. *Opt Express* 19: 20301-20315.
 21. Kim MN, Durduran T, Frangos S, Edlow BL, Buckley EM, et al. (2010) Noninvasive measurement of cerebral blood flow and blood oxygenation using near-infrared and diffuse correlation spectroscopies in critically brain-injured adults. *Neurocrit Care* 12: 173-180.
 22. Zhou C, Eucker SA, Durduran T, Yu G, Ralston J, et al. (2009) Diffuse optical monitoring of hemodynamic changes in piglet brain with closed head injury. *J Biomed Opt* 14: 034015.
 23. Yu G, Floyd TF, Durduran T, Zhou C, Wang J, et al. (2007) Validation of diffuse correlation spectroscopy for muscle blood flow with concurrent arterial spin labeled perfusion MRI. *Opt Express* 15: 1064-1075.
 24. Edlow BL, Kim MN, Durduran T, Zhou C, Putt ME, et al. (2010) The effects of healthy aging on cerebral hemodynamic responses to posture change. *Physiol Meas* 31: 477-495.
 25. Durduran T, Zhou C, Buckley EM, Kim MN, Yu G, et al. (2010) Optical measurement of cerebral hemodynamics and oxygen metabolism in neonates with congenital heart defects. *J Biomed Opt* 15: 037004.
 26. Durduran T, Zhou C, Edlow BL, Yu G, Choe R, et al. (2009) Transcranial optical monitoring of cerebrovascular hemodynamics in acute stroke patients. *Opt Express* 17: 3884-3902.
 27. Durduran T, Yu G, Burnett MG, Detre JA, Greenberg JH, et al. (2004) Diffuse optical measurement of blood flow, blood oxygenation, and metabolism in a human brain during sensorimotor cortex activation. *Opt Lett* 29: 1766-1768.
 28. Shang Y, Cheng R, Dong L, Ryan SJ, Saha SP, et al. (2011) Cerebral monitoring during carotid endarterectomy using near-infrared diffuse optical spectroscopies and electroencephalogram. *Phys Med Biol* 56: 3015-3032.
 29. Cheng R, Shang Y, Hayes D Jr, Saha SP, Yu G (2012) Noninvasive optical evaluation of spontaneous low frequency oscillations in cerebral hemodynamics. *Neuroimage* 62: 1445-1454.
 30. Dong L, Kudrimoti M, Cheng R, Shang Y, Johnson EL, et al. (2012) Noninvasive diffuse optical monitoring of head and neck tumor blood flow and oxygenation during radiation delivery. *Biomed Opt Express* 3: 259-272.
 31. Zhou C, Choe R, Shah N, Durduran T, Yu G, et al. (2007) Diffuse optical monitoring of blood flow and oxygenation in human breast cancer during early stages of neoadjuvant chemotherapy. *J Biomed Opt* 12: 051903.
 32. Sunar U, Quon H, Durduran T, Zhang J, Du J, et al. (2006) Noninvasive diffuse optical measurement of blood flow and blood oxygenation for monitoring radiation therapy in patients with head and neck tumors: a pilot study. *J Biomed Opt* 11: 064021.
 33. Durduran T, Choe R, Yu G, Zhou C, Tchou JC, et al. (2005) Diffuse optical measurement of blood flow in breast tumors. *Opt Lett* 30: 2915-2917.
 34. Shang Y, Symons TB, Durduran T, Yodh AG, Yu G (2010) Effects of muscle fiber motion on diffuse correlation spectroscopy blood flow measurements during exercise. *Biomed Opt Express* 1: 500-511.
 35. Shang Y, Zhao Y, Cheng R, Dong L, Irwin D, et al. (2009) Portable optical tissue flow oximeter based on diffuse correlation spectroscopy. *Opt Lett* 34: 3556-3558.
 36. He L, Lin Y, Shang Y, Shelton BJ, Yu G (2013) Using optical fibers with different modes to improve the signal-to-noise ratio of diffuse correlation spectroscopy flow-oximeter measurements. *J Biomed Opt* 18: 037001.
 37. Li T, Lin Y, Shang Y, He L, Huang C, et al. (2013) Simultaneous measurement of deep tissue blood flow and oxygenation using noncontact diffuse correlation spectroscopy flow-oximeter. *Sci Rep* 3: 1358.
 38. Munk N, Symons B, Shang Y, Cheng R, Yu G (2012) Noninvasively measuring the hemodynamic effects of massage on skeletal muscle: A novel hybrid near-infrared diffuse optical instrument. *J Bodyw Mov Ther* 16: 22-28.
 39. Shang Y, Gurley K, Symons B, Long D, Srikeya R, et al. (2012) Noninvasive optical characterization of muscle blood flow, oxygenation, and metabolism in women with fibromyalgia. *Arthritis Res Ther* 14: R236.
 40. Belau M, Ninck M, Hering G, Spinelli L, Contini D, et al. (2010) Noninvasive observation of skeletal muscle contraction using near-infrared time-resolved reflectance and diffusing-wave spectroscopy. *J Biomed Opt* 15: 057007.
 41. Dong J, Bi R, Ho JH, Thong PS, Soo KC, et al. (2012) Diffuse correlation spectroscopy with a fast Fourier transform-based software autocorrelator. *J Biomed Opt* 17: 07004-97001.
 42. Gurley K, Shang Y, Yu G (2012) Noninvasive optical quantification of absolute blood flow, blood oxygenation, and oxygen consumption rate in exercising skeletal muscle. *J Biomed Opt* 17: 075010.
 43. Durduran T, Yodh AG (2013) Diffuse correlation spectroscopy for non-invasive, micro-vascular cerebral blood flow measurement. *Neuroimage*.
 44. Yu G (2012) Near-infrared diffuse correlation spectroscopy in cancer diagnosis and therapy monitoring. *J Biomed Opt* 17: 010901.
 45. Mesquita RC, Durduran T, Yu G, Buckley EM, Kim MN, et al. (2011) Direct measurement of tissue blood flow and metabolism with diffuse optics. *Philos Trans A Math Phys Eng Sci* 369: 4390-4406.
 46. Durduran T, Choe R, Baker WB, Yodh AG (2010) Diffuse optics for tissue monitoring and tomography. *Rep Prog Phys* 73: 076701.
 47. Tanaka T, Riva C, Ben-Sira B (1974) Blood velocity measurements in human retinal vessels. *Science* 186: 830-831.
 48. Stern MD (1975) In vivo evaluation of microcirculation by coherent light scattering. *Nature* 254: 56-58.
 49. Brown W (1993) *Dynamic Light Scattering: The Method and Some Applications*. New York: Clarendon.
 50. Maret G, Wolf PE (1987) Multiple Light-Scattering from Disordered Media - the Effect of Brownian-Motion of Scatterers. *Zeitschrift Fur Physik B-Condensed Matter* 65: 409-413.
 51. Pine DJ, Weitz DA, Chaikin PM, Herbolzheimer E (1988) Diffusing wave spectroscopy. *Phys Rev Lett* 60: 1134-1137.
 52. Cheung C, Culver JP, Takahashi K, Greenberg JH, Yodh AG (2001) In vivo cerebrovascular measurement combining diffuse near-infrared absorption and correlation spectroscopies. *Phys Med Biol* 46: 2053-2065.
 53. Lin Y, He L, Shang Y, Yu G (2012) Noncontact diffuse correlation spectroscopy for noninvasive deep tissue blood flow measurement. *J Biomed Opt* 17: 010502.
 54. Gurley K (2012) Use of hybrid diffuse optical spectroscopies in continuous monitoring of blood flow, blood oxygenation, and oxygen consumption rate in exercising skeletal muscle. Master dissertation 1-81, University of Kentucky, Lexington, Kentucky, USA.
 55. Zelis R, Longhurst J, Capone RJ, Mason DT (1974) A comparison of regional blood flow and oxygen utilization during dynamic forearm exercise in normal subjects and patients with congestive heart failure. *Circulation* 50: 137-143.
 56. Arnold JM, Ribeiro JP, Colucci WS (1990) Muscle blood flow during forearm exercise in patients with severe heart failure. *Circulation* 82: 465-472.
 57. Habazettl H, Athanasiopoulos D, Kuebler WM, Wagner H, Roussos C, et al. (2010) Near-infrared spectroscopy and indocyanine green derived blood flow index for noninvasive measurement of muscle perfusion during exercise. *J Appl Physiol* (1985) 108: 962-967.
 58. Shang Y, Lin Y, Henry BA, Cheng R, Huang C, et al. (2013) Noninvasive evaluation of electrical stimulation impacts on muscle hemodynamics via integrating diffuse optical spectroscopies with muscle stimulator. *J Biomed Opt* 18: 105002.
 59. Irwin D, Dong L, Shang Y, Cheng R, Kudrimoti M, et al. (2011) Influences of tissue absorption and scattering on diffuse correlation spectroscopy blood flow measurements. *Biomed Opt Express* 2: 1969-1985.
 60. Dong L, He L, Lin Y, Shang Y, Yu G (2013) Simultaneously extracting multiple parameters via fitting one single autocorrelation function curve in diffuse correlation spectroscopy. *IEEE Trans Biomed Eng* 60: 361-368.