

Analysis of silage fermentation characteristics using transreflectance measurements by near infrared spectroscopy

A. Martínez, A. Soldado, R. García, D. Sánchez and B. de la Roza-Delgado

Servicio Regional de Investigación y Desarrollo Agroalimentario (SERIDA), PO Box 13, 33300 Villaviciosa (Asturias), Spain Email: admartinez@serida.org

Keywords: silage juice, fermentation parameters, FT-NIR

Introduction The fermentation end products as lactic acid, volatile fatty acids and ammonia-nitrogen, are important indicators of the efficiency of silage fermentation and are closely related to nutritive value of them (Jaster, 1995). Drying is problematic in the case of silage as many fermentation products are volatile and may get lost during the drying process. This may be a reason why NIR technology is being little used for the evaluation of silage fermentation characteristics. The feasibility of using near infrared transreflectance spectroscopy to evaluate the content in fermentation end products of grass and maize fresh silage was investigated in this study.

Materials and methods One hundred and forty-seven representative grass and sixty maize silage samples were collected over a 24 month period. Each of the silage samples were pressed immediately to obtain the extract. After centrifugation (8500 rpm, 10 min) to clarify the extract, the spectra were taken directly on extract. Samples were scanned in transreflectance mode over the range 1000 to 2500 nm in a FT-NIR Spectrum One from Perkin Elmer using IdentiCheck Reflectance Accessory (ICRA) liquids sampling accessory. To develop calibration and prediction models the Quant + software was employed. Reference data were determined on the extracts for lactic acid and total volatile fatty acids by HPLC using a WATERS Alliance 2690 instrument and ammonia-N by SPECTROQUANT ammonium test, (Merck).

Results Different mathematical pre-treatment and calibration models were developed using partial least squares (PLS) regression with internal full cross-validation. The optimum equation for each fermentation parameter was selected according to a low standard error of cross validation (SECV) and a large correlation coefficient in the calibration. R^2 values were acceptable on all fermentation parameters with $R^2 > 0.70$, except for acetic acid in grass silage (Table 1). Therefore the results indicate a reliable relationship between chemical analysis and NIR predicted values determined directly on extract. According to ASTM (American Society for Testing and Materials) guidelines for the judgement of a model's validity for NIR calibration (ASTM, 1994) and Williams and Sobering, (1996), this study obtained acceptable values of RDP (ratio of the standard deviation of the reference data to the SECV) for screening purposes, except for ammonium-N in maize silage.

Table 1 Range and statistical parameters for prediction of fermentative characteristics on grass and maize silage by FT-NIR

Grass silage				Maize silage			
Parameter (g/dl)	Range	R^2	SECV	Parameter (g/dl)	Range	R^2	SECV
Lactic acid	0-8.08	0.71	0.506	Lactic acid	0.841-5.75	0.80	0.268
Acetic acid	0-1.27	0.61	0.138	Acetic acid	0-2.30	0.85	0.148
Butyric acid	0-4.15	0.78	0.379	Butyric acid	-	-	-
Ammonium-N	0-0.261	0.82	$299 * 10^{-3}$	Ammonium-N	0-0.0012	0.71	$153 * 10^{-3}$

Conclusions The results of this investigation indicated that silage fermentation characteristics can be predicted with a fair amount of accuracy by FT-NIR analysis using ICRA as liquids sampling accessory directly on silage extract, in order to avoid the volatilisation of fermentation end products. These results may be of great practical significance when silage is used as a roughage component in dairy rations.

References

- ASTM Designation E 1655-94 (1994). American Society for Testing and Materials, 100 Barr Harbor Dr., West Conshocken, PA 19428, Annual Book of ASTM Standards.
- Jaster, E.H. (1995). Legume and grass silage preservation. In: K.J. Moore and M.A. Peterson (eds). Post-harvest physiology and preservation of forages. CSSA Special Publication 22, Madison, USA, 91-115.
- Williams, P. & D. Sobering (1996). How do we do it? A brief summary of the methods we use in developing near infrared calibrations. In: A.M.C. Davies and P. Williams (eds). *Near Infrared Spectroscopy: The future waves*. NIR Publications, Chichester, UK, 185-188.

This work was supported by Spanish INIA infrastructure project